

Red-D-Arc Welderrentals

Calibration Certificate

Miller ProHeat 35

Unit Serial Number: LK230189

Date Tested: 06/03/14

Certified Calibration

Primary Standard						
382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VCD@RC9		
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.787	383	6.014	883	9.242	1382
TC2	2.790	383	6.024	882	9.247	1382
TC3	2.791	383	6.024	882	9.242	1382
TC4	2.790	382	6.018	882	9.250	1382
TC5	2.788	382	6.009	882	9.240	1382
TC6	2.786	382	6.014	882	9.243	1382
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 08/09/2013

Precision Digital Voltage Meter SN: 2142011

Calibration Date: 08/09/2013

Machine re-calibration due date: 06/03/15

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician:

Gino Rodin Date: *06/03/14*

Signature

Red-D-Arc Welderrentals

Calibration Certificate

Miller ProHeat 35

Unit Serial Number: LK230189

Date Tested: 06/03/14

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VCD@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.787	383	6.014	883	9.242	1382
TC2	2.790	383	6.024	882	9.247	1382
TC3	2.791	383	6.024	882	9.242	1382
TC4	2.790	382	6.018	882	9.250	1382
TC5	2.788	382	6.009	882	9.240	1382
TC6	2.786	382	6.014	882	9.243	1382
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 08/09/2013

Precision Digital Voltage Meter SN: 2142011

Calibration Date: 08/09/2013

Machine re-calibration due date: 06/03/15

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician:

Gino Rodin

Date: 06/03/14

Signature